

-----  
Midmark M11 - v1.0.0  
Total Cycles: 1475

Name: RITTER 1  
Sterilizer ID: V2178198  
12/20/2021 11:35 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 179.6 F 0.0  
2:00 180.2 F 4.1  
4:00 196.4 F 9.0  
6:00 225.9 F 7.3  
8:00 228.4 F 9.8  
10:00 230.4 F 7.5  
12:00 244.5 F 13.2  
14:00 244.1 F 12.1  
16:00 257.5 F 19.1  
18:00 269.1 F 26.4

STERILIZING  
mm:ss Degrees PSI  
0:00 271.4 F 28.4  
1:00 272.4 F 28.6  
2:00 272.9 F 29.0  
3:00 271.5 F 28.5  
4:00 272.5 F 28.7  
  
Min 271.4 F 28.3  
Max 273.0 F 29.1

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 209.3 F 0.0  
5:00 225.5 F 0.0  
10:00 215.5 F 0.0  
15:00 223.0 F 0.0  
20:00 221.7 F 0.0  
25:00 224.7 F 0.0  
30:00 223.2 F 0.0  
Drying Complete

FILLING: 1:08  
HEATING: 19:43  
STERILIZING: 4:00  
VENTING: 2:59  
DRYING: 30:00  
TOTAL CYCLE: 00:57:50  
CYCLE COMPLETE

12/20/2021 01:25 PM  
ACCEPTED: X  
REJECTED:  
Operator: U02