
Midmark M11 - v1.0.5
Total Cycles: 2231

Name: RITTER2
Sterilizer ID: V2313228
04/11/2023 12:34 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 181.5 F 0.0
2:00 176.9 F 3.8
4:00 187.1 F 8.2
6:00 210.1 F 12.8
8:00 224.9 F 9.3
10:00 233.2 F 9.2
12:00 241.5 F 12.9
14:00 241.3 F 11.7
16:00 254.3 F 18.4
18:00 265.6 F 25.3

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 29.2
1:00 272.4 F 30.5
2:00 272.1 F 30.2
3:00 272.1 F 29.8
4:00 272.7 F 30.2

Min 270.8 F 29.1
Max 273.0 F 30.5

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 202.7 F 0.0
5:00 224.3 F 0.0
10:00 233.1 F 0.0
15:00 229.3 F 0.0
20:00 225.7 F 0.0
25:00 229.8 F 0.0
30:00 229.8 F 0.0
Drying Complete

FILLING: 1:34
HEATING: 19:36
STERILIZING: 4:00
VENTING: 3:13
DRYING: 30:00
TOTAL CYCLE: 00:58:23
CYCLE COMPLETE

04/11/2023 01:32 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____