
Midmark M11 - v1.0.5
Total Cycles: 2785

Name: RITTER2
Sterilizer ID: V2313228
08/17/2023 01:33 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 166.7 F 0.0
2:00 168.8 F 0.0
4:00 192.5 F 0.0
6:00 208.6 F 1.2
8:00 224.3 F 5.9
10:00 236.0 F 10.6
12:00 234.4 F 8.3
14:00 241.5 F 11.1
16:00 244.2 F 12.1
18:00 247.3 F 14.1
20:00 260.1 F 21.2
22:00 271.5 F 28.6

STERILIZING
mm:ss Degrees PSI
0:00 271.6 F 28.7
1:00 272.4 F 29.1
2:00 272.1 F 29.2
3:00 272.1 F 28.9
4:00 272.2 F 29.3

Min 271.5 F 28.5
Max 272.8 F 29.5

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 191.2 F 0.0
5:00 205.6 F 0.0
10:00 212.4 F 0.0
15:00 209.7 F 0.0
20:00 207.6 F 0.0
25:00 207.1 F 0.0
30:00 207.7 F 0.0
Drying Complete

FILLING: 1:01
HEATING: 22:37
STERILIZING: 4:00
VENTING: 3:22
DRYING: 30:00
TOTAL CYCLE: 01:00:59
CYCLE COMPLETE

08/17/2023 02:33 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____