

-----  
Midmark M11 - v1.0.5  
Total Cycles: 1496

Name: RITTER2  
Sterilizer ID: V2313228  
09/15/2022 10:03 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 198.6 F 0.0  
2:00 195.6 F 3.9  
4:00 200.5 F 9.5  
6:00 217.9 F 6.5  
8:00 232.5 F 10.8  
10:00 233.5 F 8.6  
12:00 239.6 F 10.1  
14:00 248.4 F 15.0  
16:00 261.0 F 22.0  
18:00 272.3 F 29.6

STERILIZING  
mm:ss Degrees PSI  
0:00 271.8 F 29.0  
1:00 272.8 F 29.6  
2:00 271.8 F 29.1  
3:00 272.8 F 29.6  
4:00 272.1 F 28.9  
  
Min 271.8 F 28.9  
Max 273.0 F 29.9

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 208.6 F 0.0  
5:00 214.7 F 0.0  
10:00 221.9 F 0.0  
15:00 224.3 F 0.0  
20:00 222.2 F 0.0  
25:00 219.3 F 0.0  
30:00 221.9 F 0.0  
Drying Complete

FILLING: 2:44  
HEATING: 18:29  
STERILIZING: 4:00  
VENTING: 3:07  
DRYING: 30:00  
TOTAL CYCLE: 00:58:20  
CYCLE COMPLETE

09/15/2022 11:01 AM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_