

-----  
Midmark M11 - v1.0.5  
Total Cycles: 1748

Name: RITTER2  
Sterilizer ID: V2313228  
11/21/2022 08:31 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 74.4 F 0.0  
2:00 75.8 F 0.0  
4:00 100.5 F 0.0  
6:00 114.4 F 0.0  
8:00 147.7 F 0.5  
10:00 178.7 F 1.0  
12:00 203.5 F 4.2  
14:00 215.8 F 7.9  
16:00 227.2 F 12.0  
18:00 228.9 F 7.7  
20:00 239.7 F 12.9  
22:00 239.8 F 11.0  
24:00 238.2 F 9.8  
26:00 252.1 F 16.6  
28:00 263.7 F 23.4  
30:00 271.7 F 28.7

STERILIZING  
mm:ss Degrees PSI  
0:00 271.3 F 28.9  
1:00 272.3 F 29.3  
2:00 271.8 F 29.3  
3:00 272.0 F 29.1  
4:00 272.0 F 29.5

Min 271.3 F 28.7  
Max 272.6 F 29.6

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 184.6 F 0.0  
5:00 198.4 F 0.0  
10:00 212.9 F 0.0  
15:00 222.6 F 0.0  
20:00 223.9 F 0.0  
25:00 222.6 F 0.0  
30:00 229.1 F 0.0  
Drying Complete

FILLING: 0:51  
HEATING: 30:07  
STERILIZING: 4:00  
VENTING: 3:07  
DRYING: 30:00  
TOTAL CYCLE: 01:08:05  
CYCLE COMPLETE

11/21/2022 09:39 AM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_