

-----  
Midmark M11 - v1.0.5  
Total Cycles: 3159

Name: RITTER2  
Sterilizer ID: V2313228  
11/24/2023 11:26 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 69.1 F 0.0  
2:00 73.3 F 0.0  
4:00 112.2 F 0.3  
6:00 119.0 F 2.4  
8:00 156.4 F 5.9  
10:00 182.1 F 10.4  
12:00 198.6 F 4.2  
14:00 215.5 F 7.1  
16:00 228.5 F 11.5  
18:00 230.1 F 7.5  
20:00 243.3 F 13.3  
22:00 242.4 F 11.7  
24:00 256.0 F 18.7  
26:00 268.1 F 26.3

STERILIZING  
mm:ss Degrees PSI  
0:00 271.6 F 28.7  
1:00 272.3 F 28.8  
2:00 272.8 F 29.4  
3:00 272.0 F 29.0  
4:00 272.2 F 28.7

Min 271.6 F 28.5  
Max 273.0 F 29.5

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 175.5 F 0.0  
5:00 205.5 F 0.0  
10:00 215.2 F 0.0  
15:00 217.1 F 0.0  
20:00 210.4 F 0.0  
25:00 203.2 F 0.0  
30:00 199.5 F 0.0  
Drying Complete

FILLING: 0:50  
HEATING: 27:10  
STERILIZING: 4:00  
VENTING: 2:55  
DRYING: 30:00  
TOTAL CYCLE: 01:04:55  
CYCLE COMPLETE

11/24/2023 12:30 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_