
Midmark M11 - v1.0.5
Total Cycles: 3267

Name: RITTER2
Sterilizer ID: V2313228
12/29/2023 12:58 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 89.0 F 0.0
2:00 89.9 F 0.0
4:00 97.5 F 0.6
6:00 132.2 F 3.0
8:00 167.8 F 6.6
10:00 190.4 F 10.7
12:00 217.6 F 4.9
14:00 228.6 F 8.7
16:00 237.9 F 12.5
18:00 237.7 F 10.7
20:00 235.5 F 8.9
22:00 249.7 F 15.7
24:00 261.7 F 22.4
26:00 272.7 F 29.7

STERILIZING
mm:ss Degrees PSI
0:00 271.2 F 28.9
1:00 272.2 F 29.2
2:00 272.5 F 29.8
3:00 271.1 F 29.0
4:00 272.0 F 29.3

Min 271.1 F 28.8
Max 272.7 F 29.8

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 188.3 F 0.0
5:00 231.2 F 0.0
10:00 205.1 F 0.0
15:00 208.8 F 0.0
20:00 221.0 F 0.0
25:00 229.5 F 0.0
30:00 222.6 F 0.0

Drying Complete

FILLING: 1:11
HEATING: 26:21
STERILIZING: 4:00
VENTING: 3:13
DRYING: 30:00
TOTAL CYCLE: 01:04:44
CYCLE COMPLETE

12/29/2023 02:02 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____