

-----  
Midmark M11 - v1.0.0  
Total Cycles: 2904

Name: RITTER 1  
Sterilizer ID: V2178198  
02/15/2023 08:46 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 79.9 F 0.0  
2:00 80.8 F 0.0  
4:00 97.0 F 0.0  
6:00 125.4 F 1.1  
8:00 153.6 F 4.1  
10:00 177.5 F 7.7  
12:00 195.5 F 11.5  
14:00 210.9 F 3.4  
16:00 220.2 F 7.3  
18:00 231.1 F 11.2  
20:00 228.8 F 6.8  
22:00 241.0 F 12.5  
24:00 239.4 F 10.4  
26:00 251.8 F 16.5  
28:00 262.8 F 22.9  
30:00 272.7 F 29.2

STERILIZING  
mm:ss Degrees PSI  
0:00 271.9 F 28.5  
1:00 271.8 F 29.0  
2:00 272.4 F 29.1  
3:00 271.3 F 28.4  
4:00 272.3 F 29.4

Min 271.2 F 28.4  
Max 272.7 F 29.4

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 189.6 F 0.0  
5:00 227.5 F 0.0  
10:00 212.0 F 0.0  
15:00 206.9 F 0.0  
20:00 215.2 F 0.0  
25:00 220.1 F 0.0  
30:00 224.6 F 0.0  
Drying Complete

FILLING: 0:46  
HEATING: 30:12  
STERILIZING: 4:00  
VENTING: 3:14  
DRYING: 30:00  
TOTAL CYCLE: 01:08:12  
CYCLE COMPLETE

02/15/2023 09:54 AM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_