
Midmark M11 - v1.0.0
Total Cycles: 2332

Name: RITTER 1
Sterilizer ID: V2178198
09/15/2022 12:21 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 165.7 F 0.0
2:00 167.7 F 3.7
4:00 182.5 F 8.3
6:00 204.5 F 10.3
8:00 220.7 F 8.1
10:00 234.9 F 13.4
12:00 237.1 F 10.7
14:00 237.9 F 9.0
16:00 250.4 F 15.9
18:00 262.3 F 22.7
20:00 272.6 F 29.2

STERILIZING
mm:ss Degrees PSI
0:00 271.7 F 28.7
1:00 272.8 F 29.3
2:00 271.7 F 28.6
3:00 272.8 F 29.5
4:00 272.3 F 28.9

Min 271.7 F 28.5
Max 272.9 F 29.5

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 210.6 F 0.0
5:00 183.1 F 0.0
10:00 192.1 F 0.0
15:00 191.6 F 0.0
20:00 188.3 F 0.0
25:00 190.1 F 0.0
30:00 189.5 F 0.0
Drying Complete

FILLING: 1:44
HEATING: 20:17
STERILIZING: 4:00
VENTING: 3:11
DRYING: 30:00
TOTAL CYCLE: 00:59:13
CYCLE COMPLETE

09/15/2022 01:20 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____