
Midmark M11 - v1.0.5
Total Cycles: 360

Name: RITTER2
Sterilizer ID: V2313228
09/14/2021 09:23 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 163.8 F 0.0
2:00 167.1 F 2.9
4:00 187.2 F 7.8
6:00 211.7 F 11.1
8:00 221.4 F 10.1
10:00 233.3 F 8.0
12:00 240.6 F 10.9
14:00 246.3 F 13.9
16:00 258.8 F 20.7
18:00 268.9 F 27.9

STERILIZING
mm:ss Degrees PSI
0:00 271.9 F 30.0
1:00 272.0 F 30.1
2:00 272.3 F 30.0
3:00 272.0 F 30.2
4:00 271.6 F 30.3

Min 271.2 F 29.9
Max 272.5 F 30.3

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 200.4 F 0.0
5:00 235.3 F 0.0
10:00 224.3 F 0.0
15:00 224.1 F 0.0
20:00 221.2 F 0.0
Dry Cycle Aborted

FILLING: 1:52
HEATING: 18:51
STERILIZING: 4:00
VENTING: 3:02
DRYING: 23:45
TOTAL CYCLE: 00:51:30

09/14/2021 10:24 AM
ACCEPTED: X
REJECTED:
Operator: U01