
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.4uS / 0.2ppm
CYCLE NUMBER 000191
15:14 17/09/2020
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 1:12
PRESSURIZING 1:14
CYCLE ABORTED 4:23
STOP BUTTON PRESSED
NOT STERILE
END TIME 4:23

Digital Signature #
5DD333657A05F08E
