
Midmark M11 - v1.0.0
Total Cycles: 1511

Name: RITTER 1
Sterilizer ID: V2178198
01/11/2022 01:53 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 208.1 F 0.0
2:00 202.2 F 4.6
4:00 211.2 F 9.6
6:00 225.3 F 5.6
8:00 235.5 F 10.8
10:00 236.1 F 7.9
12:00 241.6 F 10.9
14:00 246.5 F 13.4
16:00 259.6 F 20.5
18:00 271.1 F 27.9

STERILIZING
mm:ss Degrees PSI
0:00 271.3 F 28.2
1:00 272.2 F 28.7
2:00 271.6 F 28.5
3:00 272.1 F 28.6
4:00 271.4 F 28.4

Min 271.2 F 28.1
Max 272.4 F 29.1

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 215.9 F 0.0
5:00 224.6 F 0.0
10:00 236.8 F 0.0
15:00 237.7 F 0.0
20:00 235.3 F 0.0
25:00 223.9 F 0.0
30:00 215.0 F 0.0
Drying Complete

FILLING: 2:16
HEATING: 18:42
STERILIZING: 4:00
VENTING: 3:11
DRYING: 30:00
TOTAL CYCLE: 00:58:08
CYCLE COMPLETE

01/11/2022 02:51 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____