
Midmark M11 - v1.0.5
Total Cycles: 3726

Name: RITTER2
Sterilizer ID: V2313228
06/21/2024 10:16 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 113.1 F 0.0
2:00 113.2 F 0.4
4:00 123.4 F 2.1
6:00 146.0 F 4.7
8:00 170.2 F 7.9
10:00 190.2 F 11.4
12:00 209.0 F 4.2
14:00 214.4 F 7.7
16:00 226.5 F 11.6
18:00 228.0 F 7.3
20:00 239.5 F 12.3
22:00 237.9 F 10.2
24:00 249.6 F 15.7
26:00 259.6 F 21.4
28:00 269.0 F 27.5

STERILIZING
mm:ss Degrees PSI
0:00 271.3 F 29.3
1:00 272.7 F 30.0
2:00 271.8 F 29.4
3:00 271.5 F 29.7
4:00 272.6 F 29.9

Min 271.2 F 29.2
Max 272.9 F 30.1

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 191.3 F 0.0
5:00 200.6 F 0.0
10:00 194.8 F 0.0
15:00 188.9 F 0.0
20:00 191.3 F 0.0
25:00 188.2 F 0.0
30:00 186.2 F 0.0
Drying Complete

FILLING: 0:53
HEATING: 29:16
STERILIZING: 4:00
VENTING: 3:43
DRYING: 30:00
TOTAL CYCLE: 01:07:51
CYCLE COMPLETE

06/21/2024 11:23 AM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____