
Midmark M11 - v1.0.5
Total Cycles: 3186

Name: RITTER2
Sterilizer ID: V2313228
12/05/2023 02:50 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 95.8 F 0.0
2:00 97.0 F 0.2
4:00 126.0 F 1.9
6:00 150.7 F 4.6
8:00 175.1 F 7.7
10:00 204.2 F 11.0
12:00 210.0 F 4.9
14:00 220.5 F 8.5
16:00 231.9 F 12.9
18:00 233.2 F 9.5
20:00 238.0 F 10.0
22:00 244.9 F 13.6
24:00 256.7 F 19.8
26:00 267.3 F 26.5

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 29.4
1:00 272.6 F 30.2
2:00 271.9 F 29.6
3:00 271.4 F 29.7
4:00 272.6 F 30.2

Min 271.2 F 29.3
Max 272.8 F 30.3

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 202.7 F 0.0
5:00 197.8 F 0.0
10:00 192.1 F 0.0
15:00 185.7 F 0.0
20:00 185.6 F 0.0
25:00 184.4 F 0.0
30:00 182.8 F 0.0
Drying Complete

FILLING: 0:54
HEATING: 27:30
STERILIZING: 4:00
VENTING: 3:38
DRYING: 30:00
TOTAL CYCLE: 01:06:02
CYCLE COMPLETE

12/05/2023 03:55 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____