
Midmark M11 - v1.0.5
Total Cycles: 949

Name: RITTER2
Sterilizer ID: V2313228
04/04/2022 08:52 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 109.9 F 0.0
2:00 112.2 F 0.0
4:00 125.8 F 0.0
6:00 164.1 F 0.7
8:00 187.7 F 4.3
10:00 208.4 F 8.1
12:00 216.7 F 12.8
14:00 225.1 F 8.1
16:00 237.7 F 13.3
18:00 238.8 F 11.1
20:00 238.1 F 9.9
22:00 251.8 F 16.7
24:00 263.4 F 23.4
26:00 272.2 F 29.0

STERILIZING
mm:ss Degrees PSI
0:00 272.1 F 28.9
1:00 271.7 F 29.2
2:00 272.7 F 29.6
3:00 271.8 F 28.9
4:00 271.7 F 29.3

Min 271.4 F 28.7
Max 272.8 F 29.8

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 192.1 F 0.0
5:00 202.6 F 0.0
10:00 199.7 F 0.0
15:00 203.4 F 0.0
20:00 210.6 F 0.0
25:00 218.8 F 0.0
30:00 220.4 F 0.0
Drying Complete

FILLING: 0:51
HEATING: 26:02
STERILIZING: 4:00
VENTING: 3:10
DRYING: 30:00
TOTAL CYCLE: 01:04:03
CYCLE COMPLETE

04/04/2022 10:19 AM
ACCEPTED: X
REJECTED:
Operator: U02