
Midmark M11 - v1.0.5
Total Cycles: 467

Name: RITTER2
Sterilizer ID: V2313228
10/21/2021 09:12 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 199.8 F 0.0
2:00 193.7 F 3.2
4:00 203.9 F 8.6
6:00 220.2 F 8.9
8:00 231.4 F 10.4
10:00 232.8 F 8.0
12:00 241.7 F 11.4
14:00 246.2 F 13.7
16:00 258.7 F 20.5
18:00 269.8 F 27.7

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 28.8
1:00 272.5 F 29.4
2:00 271.7 F 29.1
3:00 272.3 F 29.3
4:00 271.5 F 28.9

Min 271.5 F 28.7
Max 272.7 F 29.7

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 204.7 F 0.0
5:00 236.9 F 0.0
10:00 227.0 F 0.0
15:00 227.8 F 0.0
20:00 222.6 F 0.0
25:00 217.1 F 0.0
30:00 212.7 F 0.0

Drying Complete

FILLING: 2:26
HEATING: 18:55
STERILIZING: 4:00
VENTING: 3:03
DRYING: 30:00
TOTAL CYCLE: 00:58:23
CYCLE COMPLETE

10/21/2021 10:11 AM
ACCEPTED: X
REJECTED:
Operator: U02