
Midmark M11 - v1.0.5
Total Cycles: 2013

Name: RITTER2
Sterilizer ID: V2313228
02/06/2023 03:46 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 146.4 F 0.0
2:00 145.7 F 0.0
4:00 182.6 F 1.6
6:00 201.1 F 6.3
8:00 223.1 F 11.6
10:00 228.6 F 7.3
12:00 239.7 F 12.9
14:00 240.0 F 11.2
16:00 239.2 F 10.2
18:00 253.3 F 17.3
20:00 265.3 F 24.6

STERILIZING
mm:ss Degrees PSI
0:00 271.2 F 28.8
1:00 272.5 F 29.5
2:00 271.5 F 28.7
3:00 272.5 F 29.7
4:00 272.1 F 29.1

Min 271.2 F 28.7
Max 272.6 F 29.7

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 207.4 F 0.0
5:00 227.6 F 0.0
10:00 238.3 F 0.0
15:00 233.3 F 0.0
20:00 224.8 F 0.0
25:00 219.5 F 0.0
30:00 218.9 F 0.0
Drying Complete

FILLING: 1:02
HEATING: 21:44
STERILIZING: 4:00
VENTING: 3:08
DRYING: 30:00
TOTAL CYCLE: 00:59:54
CYCLE COMPLETE

02/06/2023 04:45 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____