
Midmark M11 - v1.0.5
Total Cycles: 2395

Name: RITTER2
Sterilizer ID: V2313228
05/19/2023 10:42 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 147.7 F 0.0
2:00 147.2 F 2.3
4:00 159.8 F 5.4
6:00 188.5 F 10.3
8:00 209.7 F 5.3
10:00 223.7 F 9.3
12:00 232.2 F 9.7
14:00 239.7 F 11.8
16:00 239.3 F 10.4
18:00 252.7 F 17.1
20:00 264.6 F 24.3

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 29.0
1:00 272.6 F 29.5
2:00 271.5 F 29.0
3:00 272.7 F 29.6
4:00 271.8 F 28.9

Min 271.5 F 28.8
Max 272.8 F 29.8

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 207.5 F 0.0
5:00 208.8 F 0.0
10:00 218.1 F 0.0
15:00 223.0 F 0.0
20:00 223.3 F 0.0
25:00 223.3 F 0.0
30:00 224.1 F 0.0
Drying Complete

FILLING: 1:06
HEATING: 21:50
STERILIZING: 4:00
VENTING: 3:12
DRYING: 30:00
TOTAL CYCLE: 01:00:08
CYCLE COMPLETE

05/19/2023 11:42 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____