
Midmark M11 - v1.0.0
Total Cycles: 2147

Name: RITTER 1
Sterilizer ID: V2178198
07/19/2022 07:04 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 74.2 F 0.0
2:00 77.0 F 0.0
4:00 107.3 F 0.0
6:00 138.4 F 0.0
8:00 159.2 F 0.6
10:00 179.8 F 3.5
12:00 199.8 F 6.8
14:00 215.0 F 9.9
16:00 225.5 F 13.2
18:00 226.0 F 6.6
20:00 235.0 F 10.5
22:00 234.9 F 8.5
24:00 242.2 F 11.8
26:00 238.8 F 9.8
28:00 251.9 F 16.2
30:00 262.7 F 22.4
32:00 272.6 F 29.1

STERILIZING
mm:ss Degrees PSI
0:00 271.1 F 28.3
1:00 272.7 F 29.1
2:00 271.5 F 28.2
3:00 272.1 F 29.1
4:00 272.5 F 28.7

Min 271.0 F 28.2
Max 272.7 F 29.2

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 180.8 F 0.0
5:00 222.7 F 0.0
10:00 220.3 F 0.0
15:00 218.7 F 0.0
20:00 212.6 F 0.0
25:00 208.1 F 0.0
30:00 200.2 F 0.0

Drying Complete

FILLING: 0:52
HEATING: 32:26
STERILIZING: 4:00
VENTING: 3:11
DRYING: 30:00
TOTAL CYCLE: 01:10:29
CYCLE COMPLETE

07/19/2022 08:14 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____