

-----  
Herzig Eye Institute Ottawa Inc  
phone/email  
150 Bloor St. West, Suite 210  
Toronto ON M5S2X9  
North York, ON M5S2X9  
Canada

6472616554  
david@herzig-eye.com  
-----

STATIM 2000 S203R751  
SN 500919I00001  
Unit # : 000  
WATER QUALITY  
0.7uS / 0.4ppm  
CYCLE NUMBER 002688  
12:17 15/11/2021  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 2:59  
PRESSURIZING 3:01  
STERILIZING 4:04  
136.0 C 220kPa 4:04  
136.4 C 225kPa 4:34  
136.4 C 224kPa 5:04  
136.2 C 218kPa 5:34  
136.2 C 223kPa 6:04  
136.4 C 224kPa 6:34  
136.4 C 223kPa 7:04  
136.5 C 225kPa 7:34  
136.5 C 225kPa 7:34  
VENTING 7:34  
AIR DRYING 8:28  
CYCLE COMPLETE 8:36  
-----

Digital Signature #  
34EC60B886C46C16  
-----