
Midmark M11 - v1.0.5
Total Cycles: 3259

Name: RITTER2
Sterilizer ID: V2313228
12/28/2023 09:40 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 77.2 F 0.0
2:00 77.7 F 0.0
4:00 85.1 F 0.3
6:00 116.9 F 1.7
8:00 153.4 F 4.6
10:00 177.5 F 7.8
12:00 196.9 F 12.0
14:00 214.7 F 5.1
16:00 222.7 F 8.9
18:00 233.0 F 13.1
20:00 234.9 F 9.2
22:00 239.4 F 10.4
24:00 244.8 F 13.1
26:00 256.7 F 19.4
28:00 267.3 F 26.1

STERILIZING
mm:ss Degrees PSI
0:00 271.1 F 28.9
1:00 272.5 F 29.4
2:00 271.3 F 29.2
3:00 272.5 F 29.3
4:00 271.2 F 29.1

Min 271.0 F 28.7
Max 272.7 F 29.7

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 177.1 F 0.0
5:00 237.8 F 0.0
10:00 196.1 F 0.0
15:00 198.6 F 0.0
20:00 181.9 F 0.0
25:00 183.6 F 0.0
30:00 186.7 F 0.0
Drying Complete

FILLING: 0:51
HEATING: 29:25
STERILIZING: 4:00
VENTING: 3:11
DRYING: 30:00
TOTAL CYCLE: 01:07:27
CYCLE COMPLETE

12/28/2023 10:47 AM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____