
Midmark M11 - v1.0.0
Total Cycles: 3979

Name: RITTER 1
Sterilizer ID: V2178198
11/03/2023 01:28 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 94.8 F 0.0
2:00 120.9 F 1.0
4:00 153.4 F 3.3
6:00 173.2 F 6.7
8:00 183.8 F 10.7
10:00 211.5 F 3.7
12:00 222.9 F 6.6
14:00 232.0 F 10.8
16:00 232.5 F 6.9
18:00 240.8 F 11.9
20:00 238.8 F 9.8
22:00 251.7 F 16.1
24:00 263.2 F 22.8
26:00 271.9 F 28.3

STERILIZING
mm:ss Degrees PSI
0:00 271.8 F 28.2
1:00 272.6 F 29.1
2:00 271.8 F 28.2
3:00 272.6 F 29.1
4:00 272.1 F 28.4

Min 271.4 F 28.1
Max 272.7 F 29.1

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 206.7 F 0.0
5:00 205.2 F 0.0
10:00 222.2 F 0.0
15:00 225.3 F 0.0
20:00 224.1 F 0.0
25:00 224.7 F 0.0
30:00 225.5 F 0.0
Drying Complete

FILLING: 0:60
HEATING: 26:01
STERILIZING: 4:00
VENTING: 3:23
DRYING: 30:00
TOTAL CYCLE: 01:04:24
CYCLE COMPLETE

11/03/2023 02:32 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____