
Midmark M11 - v1.0.0
Total Cycles: 5013

Name: RITTER 1
Sterilizer ID: V2178198
10/22/2024 08:35 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 68.6 F 0.0
2:00 73.2 F 0.0
4:00 112.3 F 0.0
6:00 120.8 F 0.1
8:00 152.1 F 0.4
10:00 177.9 F 0.9
12:00 197.4 F 3.9
14:00 207.5 F 6.9
16:00 219.1 F 10.2
18:00 226.3 F 10.1
20:00 230.2 F 8.3
22:00 241.1 F 12.9
24:00 238.7 F 10.0
26:00 237.5 F 8.6
28:00 248.6 F 14.2
30:00 259.7 F 20.3
32:00 270.0 F 26.8

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 27.9
1:00 272.6 F 28.5
2:00 271.4 F 27.8
3:00 272.6 F 28.6
4:00 271.7 F 27.9

Min 271.4 F 27.8
Max 272.6 F 28.7

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 200.7 F 0.0
5:00 198.0 F 0.0
10:00 205.1 F 0.0
15:00 210.9 F 0.0
20:00 211.2 F 0.0
25:00 214.6 F 0.0
30:00 218.2 F 0.0

Drying Complete

FILLING: 0:42
HEATING: 32:58
STERILIZING: 4:00
VENTING: 3:25
DRYING: 30:00
TOTAL CYCLE: 01:11:05
CYCLE COMPLETE

10/22/2024 09:46 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____