
Midmark M11 - v1.0.5
Total Cycles: 1513

Name: RITTER2
Sterilizer ID: V2313228
09/21/2022 10:05 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 175.3 F 0.0
2:00 176.8 F 3.4
4:00 180.7 F 7.6
6:00 201.9 F 12.8
8:00 223.4 F 7.6
10:00 233.2 F 13.2
12:00 237.3 F 10.7
14:00 236.3 F 9.4
16:00 251.0 F 16.5
18:00 263.3 F 23.7
20:00 271.7 F 29.1

STERILIZING
mm:ss Degrees PSI
0:00 271.2 F 29.2
1:00 272.4 F 30.0
2:00 272.1 F 29.5
3:00 271.2 F 29.1
4:00 271.9 F 30.0

Min 271.1 F 29.1
Max 272.5 F 30.1

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 202.8 F 0.0
5:00 212.1 F 0.0
10:00 230.5 F 0.0
15:00 233.6 F 0.0
20:00 234.3 F 0.0
25:00 223.9 F 0.0
30:00 226.6 F 0.0
Drying Complete

FILLING: 1:10
HEATING: 20:08
STERILIZING: 4:00
VENTING: 3:10
DRYING: 30:00
TOTAL CYCLE: 00:58:28
CYCLE COMPLETE

09/21/2022 11:03 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____