

-----  
North York, ON  
Canada  
-----

STATIM 2000 S203R750  
SN 500919I00007  
Unit # : 000  
WATER QUALITY  
1.1uS / 0.7ppm  
CYCLE NUMBER 001096  
8:48 26/05/2022  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 3:58  
PRESSURIZING 4:03  
STERILIZING 5:43  
136.5 C 224kPa 5:43  
136.2 C 223kPa 6:13  
136.3 C 223kPa 6:43  
136.1 C 223kPa 7:13  
136.4 C 224kPa 7:43  
136.3 C 224kPa 8:13  
136.4 C 225kPa 8:43  
136.2 C 223kPa 9:13  
136.2 C 223kPa 9:13  
VENTING 9:13  
AIR DRYING 10:08  
CYCLE COMPLETE 25:09  
-----

Digital Signature #  
843DD9283E843C59  
-----