

-----  
Midmark M11 - v1.0.0  
Total Cycles: 1325

Name: RITTER 1  
Sterilizer ID: V2178198  
11/01/2021 08:23 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 146.4 F 0.0  
2:00 151.6 F 0.7  
4:00 159.4 F 3.1  
6:00 178.4 F 6.1  
8:00 199.7 F 10.2  
10:00 215.1 F 6.0  
12:00 223.6 F 8.1  
14:00 235.4 F 12.7  
16:00 236.5 F 9.6  
18:00 239.9 F 9.3  
20:00 247.1 F 13.7  
22:00 258.3 F 20.0  
24:00 268.9 F 26.6

STERILIZING  
mm:ss Degrees PSI  
0:00 271.5 F 28.5  
1:00 272.7 F 29.1  
2:00 271.6 F 28.6  
3:00 272.5 F 29.0  
4:00 271.4 F 28.4

Min 271.4 F 28.4  
Max 272.8 F 29.3

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 212.9 F 0.0  
5:00 211.9 F 0.0  
10:00 214.3 F 0.0  
15:00 220.2 F 0.0  
20:00 217.8 F 0.0  
25:00 221.8 F 0.0  
30:00 222.8 F 0.0  
Drying Complete

FILLING: 0:58  
HEATING: 25:08  
STERILIZING: 4:00  
VENTING: 3:21  
DRYING: 30:00  
TOTAL CYCLE: 01:03:27  
CYCLE COMPLETE

11/01/2021 09:27 AM  
ACCEPTED: X  
REJECTED:  
Operator: U01