
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 001150
8:35 10/02/2021
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:28
PRESSURIZING 2:29
STERILIZING 3:29
136.4 C 220kPa 3:29
136.1 C 222kPa 3:59
136.0 C 221kPa 4:29
136.4 C 220kPa 4:59
136.4 C 222kPa 5:29
136.3 C 224kPa 5:59
136.1 C 222kPa 6:29
136.4 C 223kPa 6:59
136.4 C 223kPa 6:59
VENTING 6:59
AIR DRYING 7:50
CYCLE COMPLETE 7:54

Digital Signature #
F33EEDE1CF9E5135
