
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.4uS / 0.2ppm
CYCLE NUMBER 006543
14:31 01/11/2023
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:41
PRESSURIZING 3:42
STERILIZING 4:42
136.4 C 222kPa 4:42
136.1 C 220kPa 5:13
136.2 C 222kPa 5:43
136.4 C 224kPa 6:13
136.3 C 223kPa 6:43
136.2 C 218kPa 7:13
136.4 C 223kPa 7:43
136.3 C 224kPa 8:13
136.3 C 224kPa 8:13
VENTING 8:13
AIR DRYING 9:07
CYCLE COMPLETE 18:00

Digital Signature #
2452FB01DB35D8ED
