
Brampton, ON
Canada

STATIM 5000 S501R713
SN 511019D00004
Unit # : 000
WATER QUALITY
1.0uS / 0.6ppm
CYCLE NUMBER 001496
14:14 10/13/2021
UNWRAPPED
134 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 4:01
PRESSURIZING 5:58
STERILIZING 7:42
136.1 C 220kPa 7:42
136.4 C 222kPa 8:12
136.0 C 219kPa 8:42
136.3 C 222kPa 9:12
136.4 C 219kPa 9:42
135.8 C 217kPa 10:12
136.4 C 222kPa 10:42
136.4 C 223kPa 11:12
VENTING 11:12
AIR DRYING 12:15
CYCLE COMPLETE 36:52

Digital Signature #
F0D682B40F5E5885
