
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 003964
14:35 07/07/2022
RUBBER AND PLASTICS
121 C for 15 min.
CYCLE START 0:00
CONDITIONING 2:47
PRESSURIZING 4:20
STERILIZING 4:56
123.5 C 116kPa 4:56
123.4 C 118kPa 5:26
123.4 C 119kPa 5:56
123.4 C 118kPa 6:26
123.4 C 119kPa 6:56
123.4 C 119kPa 7:26
123.2 C 116kPa 7:56
123.4 C 118kPa 8:26
123.4 C 118kPa 8:56
123.4 C 119kPa 9:26
123.1 C 116kPa 9:56
123.3 C 118kPa 10:26
123.4 C 117kPa 10:56
123.2 C 118kPa 11:26
123.4 C 118kPa 11:56
123.3 C 118kPa 12:26
123.4 C 116kPa 12:56
123.4 C 118kPa 13:26
123.2 C 118kPa 13:56
123.4 C 116kPa 14:26
123.4 C 118kPa 14:56
123.4 C 118kPa 15:26
123.4 C 118kPa 15:56
123.4 C 118kPa 16:26
123.4 C 119kPa 16:56
123.4 C 118kPa 17:26

123.2 C 118kPa 17:56
123.4 C 118kPa 18:26
123.4 C 118kPa 18:56
123.4 C 118kPa 19:26
123.4 C 117kPa 19:56
123.4 C 117kPa 19:56
VENTING 19:56
AIR DRYING 20:40
CYCLE COMPLETE 26:45

Digital Signature #
A4F2A01A573E9727
