
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.0uS / 0.0ppm
CYCLE NUMBER 002384
10:32 16/12/2024
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 1:55
PRESSURIZING 2:01
STERILIZING 3:09
136.4 C 224kPa 3:09
136.3 C 224kPa 3:39
136.4 C 226kPa 4:09
136.3 C 223kPa 4:39
136.1 C 222kPa 5:09
136.1 C 223kPa 5:39
136.4 C 226kPa 6:09
136.4 C 222kPa 6:39
136.4 C 222kPa 6:39
VENTING 6:39
AIR DRYING 7:31
CYCLE COMPLETE 14:00

Digital Signature #
9F55C3EDC1BF6E0C
