

-----  
North York, ON  
Canada  
-----

STATIM 2000 S203R750  
SN 500919I00007  
Unit # : 000  
WATER QUALITY  
0.2uS / 0.1ppm  
CYCLE NUMBER 001288  
13:57 07/10/2022  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 2:48  
PRESSURIZING 2:56  
STERILIZING 4:07  
136.2 C 221kPa 4:07  
136.3 C 224kPa 4:37  
136.4 C 225kPa 5:07  
136.4 C 223kPa 5:37  
135.8 C 220kPa 6:07  
136.4 C 225kPa 6:37  
136.2 C 222kPa 7:07  
136.1 C 222kPa 7:37  
136.1 C 222kPa 7:37  
VENTING 7:37  
AIR DRYING 8:29  
CYCLE COMPLETE 23:29  
-----

Digital Signature #  
5AF9A8FC486FAF0A  
-----