
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.3uS / 0.1ppm
CYCLE NUMBER 003109
11:14 08/02/2022
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:55
PRESSURIZING 2:56
STERILIZING 3:58
136.4 C 222kPa 3:58
136.2 C 223kPa 4:28
136.2 C 222kPa 4:58
136.2 C 223kPa 5:28
136.4 C 221kPa 5:58
136.3 C 224kPa 6:28
136.2 C 222kPa 6:58
136.4 C 221kPa 7:28
136.4 C 221kPa 7:28
VENTING 7:29
AIR DRYING 8:26
CYCLE COMPLETE 8:29

Digital Signature #
57C7E3C9963298BA
