
Midmark M11 - v1.0.5
Total Cycles: 465

Name: RITTER2
Sterilizer ID: V2313228
10/21/2021 06:44 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 72.7 F 0.0
2:00 73.6 F 0.0
4:00 77.8 F 0.0
6:00 91.6 F 0.1
8:00 141.2 F 0.2
10:00 169.1 F 1.0
12:00 185.1 F 4.6
14:00 203.9 F 7.9
16:00 216.2 F 11.4
18:00 222.7 F 5.9
20:00 231.9 F 10.3
22:00 235.7 F 9.3
24:00 242.2 F 12.8
26:00 241.3 F 11.6
28:00 253.5 F 17.8
30:00 264.6 F 24.5

STERILIZING
mm:ss Degrees PSI
0:00 271.7 F 29.2
1:00 272.7 F 29.7
2:00 272.0 F 29.4
3:00 272.7 F 29.7
4:00 271.9 F 29.2

Min 271.7 F 29.0
Max 273.0 F 30.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 178.0 F 0.0
5:00 234.3 F 0.0
10:00 216.1 F 0.0
15:00 188.4 F 0.0
20:00 192.5 F 0.0
25:00 193.6 F 0.0
30:00 197.4 F 0.0
Drying Complete

FILLING: 0:56
HEATING: 31:57
STERILIZING: 4:00
VENTING: 3:07
DRYING: 30:00
TOTAL CYCLE: 01:09:59
CYCLE COMPLETE

10/21/2021 07:53 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____