

-----  
Midmark M11 - v1.0.0  
Total Cycles: 2991

Name: RITTER 1  
Sterilizer ID: V2178198  
03/09/2023 03:32 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 182.4 F 0.0  
2:00 179.5 F 4.0  
4:00 196.5 F 8.9  
6:00 219.4 F 8.0  
8:00 232.4 F 9.4  
10:00 238.4 F 8.8  
12:00 243.6 F 12.9  
14:00 243.2 F 11.8  
16:00 256.9 F 19.0  
18:00 268.8 F 26.5

STERILIZING  
mm:ss Degrees PSI  
0:00 271.4 F 28.3  
1:00 272.4 F 28.9  
2:00 271.3 F 28.3  
3:00 272.5 F 29.1  
4:00 271.8 F 28.4  
  
Min 271.2 F 28.2  
Max 272.7 F 29.2

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 212.7 F 0.0  
5:00 237.0 F 0.0  
10:00 227.9 F 0.0  
15:00 228.3 F 0.0  
20:00 231.5 F 0.0  
25:00 226.6 F 0.0  
30:00 228.9 F 0.0  
Drying Complete

FILLING: 1:38  
HEATING: 19:05  
STERILIZING: 4:00  
VENTING: 3:14  
DRYING: 30:00  
TOTAL CYCLE: 00:57:58  
CYCLE COMPLETE

03/09/2023 04:29 PM

ACCEPTED:

REJECTED:

Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_