
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.3uS / 0.1ppm
CYCLE NUMBER 001265
15:40 21/09/2022
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:18
PRESSURIZING 3:24
STERILIZING 4:44
136.0 C 220kPa 4:44
135.9 C 222kPa 5:14
136.4 C 225kPa 5:44
136.2 C 224kPa 6:14
136.4 C 225kPa 6:44
136.4 C 223kPa 7:14
136.4 C 225kPa 7:44
136.4 C 224kPa 8:14
136.4 C 224kPa 8:14
VENTING 8:14
CYCLE ABORTED 8:53
STOP BUTTON PRESSED
NOT STERILE
END TIME 8:53

Digital Signature #
8B98BA9269AF9F8A
