
Midmark M11 - v1.0.5
Total Cycles: 1326

Name: RITTER2
Sterilizer ID: V2313228
07/20/2022 02:41 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 184.4 F 0.0
2:00 181.4 F 0.2
4:00 194.4 F 3.2
6:00 219.2 F 8.1
8:00 233.2 F 12.0
10:00 236.1 F 10.6
12:00 234.3 F 8.8
14:00 238.9 F 10.0
16:00 248.9 F 15.1
18:00 261.0 F 22.0
20:00 272.1 F 29.3

STERILIZING
mm:ss Degrees PSI
0:00 271.3 F 29.0
1:00 272.2 F 29.3
2:00 271.6 F 29.4
3:00 272.1 F 29.2
4:00 271.6 F 29.3

Min 271.3 F 28.8
Max 272.5 F 29.7

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 204.9 F 0.0
5:00 228.5 F 0.0
10:00 235.7 F 0.0
15:00 232.7 F 0.0
20:00 228.6 F 0.0
25:00 230.8 F 0.0
30:00 229.2 F 0.0
Drying Complete

FILLING: 1:06
HEATING: 20:30
STERILIZING: 4:00
VENTING: 3:13
DRYING: 30:00
TOTAL CYCLE: 00:58:49
CYCLE COMPLETE

07/20/2022 03:39 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____