

-----  
Midmark M11 - v1.0.5  
Total Cycles: 3913

Name: RITTER2  
Sterilizer ID: V2313228  
09/06/2024 01:54 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 108.3 F 0.0  
2:00 109.4 F 0.0  
4:00 115.6 F 0.3  
6:00 149.1 F 3.7  
8:00 180.9 F 7.6  
10:00 222.2 F 12.3  
12:00 220.4 F 6.7  
14:00 233.0 F 12.0  
16:00 235.7 F 9.0  
18:00 238.8 F 10.3  
20:00 246.4 F 13.7  
22:00 258.5 F 20.3  
24:00 269.4 F 27.3

STERILIZING  
mm:ss Degrees PSI  
0:00 271.3 F 28.9  
1:00 271.8 F 28.8  
2:00 271.8 F 29.2  
3:00 271.8 F 29.5  
4:00 271.1 F 29.2

Min 271.0 F 28.7  
Max 272.0 F 29.6

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 190.5 F 0.0  
5:00 214.8 F 0.0  
10:00 211.4 F 0.0  
15:00 210.8 F 0.0  
20:00 209.9 F 0.0  
25:00 211.0 F 0.0  
30:00 210.8 F 0.0  
Drying Complete

FILLING: 0:58  
HEATING: 24:57  
STERILIZING: 4:00  
VENTING: 3:18  
DRYING: 30:00  
TOTAL CYCLE: 01:03:13  
CYCLE COMPLETE

09/06/2024 02:57 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_