
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.1uS / 0.0ppm
CYCLE NUMBER 006915
7:28 08/02/2024

UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:00
PRESSURIZING 3:01
STERILIZING 4:02
136.4 C 221kPa 4:02
136.4 C 220kPa 4:32
136.0 C 222kPa 5:02
136.4 C 217kPa 5:32
136.1 C 222kPa 6:02
136.2 C 221kPa 6:32
135.8 C 215kPa 7:02
136.4 C 220kPa 7:32
136.4 C 220kPa 7:32
VENTING 7:32
AIR DRYING 8:24
CYCLE COMPLETE 8:34

Digital Signature #
CE1060B8995D54F5
