

-----  
Herzig Eye Institute Ottawa Inc  
phone/email  
150 Bloor St. West, Suite 210  
Toronto ON M5S2X9  
North York, ON M5S2X9  
Canada

6472616554  
david@herzig-eye.com  
-----

STATIM 2000 S203R751  
SN 500919I00001  
Unit # : 000  
WATER QUALITY  
0.2uS / 0.1ppm  
CYCLE NUMBER 005739  
11:09 12/06/2023  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 1:56  
PRESSURIZING 2:00  
STERILIZING 2:58  
136.4 C 222kPa 2:58  
136.4 C 225kPa 3:28  
136.4 C 225kPa 3:58  
136.2 C 223kPa 4:28  
136.5 C 225kPa 4:58  
136.4 C 225kPa 5:28  
136.2 C 221kPa 5:58  
136.3 C 225kPa 6:28  
136.3 C 225kPa 6:28  
VENTING 6:28  
AIR DRYING 7:17  
CYCLE COMPLETE 7:40  
-----

Digital Signature #  
869CC7BB437E284C  
-----