
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.0uS / 0.0ppm
CYCLE NUMBER 006059
8:12 02/08/2023
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:30
PRESSURIZING 2:33
STERILIZING 3:22
136.0 C 220kPa 3:22
136.4 C 224kPa 3:52
136.3 C 221kPa 4:22
136.4 C 222kPa 4:52
136.4 C 223kPa 5:22
136.3 C 221kPa 5:52
136.0 C 219kPa 6:22
136.2 C 221kPa 6:52
136.2 C 221kPa 6:52
VENTING 6:52
AIR DRYING 7:42
CYCLE COMPLETE 10:05

Digital Signature #
4D16DDD51EB857EC
