
Midmark M11 - v1.0.5
Total Cycles: 4435

Name: RITTER2
Sterilizer ID: V2313228
03/19/2025 01:46 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 89.3 F 0.0
2:00 90.8 F 0.4
4:00 93.8 F 1.5
6:00 135.3 F 4.1
8:00 160.4 F 6.9
10:00 181.6 F 10.0
12:00 204.2 F 10.8
14:00 214.0 F 6.5
16:00 222.8 F 10.4
18:00 232.9 F 8.5
20:00 235.8 F 10.9
22:00 235.2 F 8.2
24:00 246.5 F 14.3
26:00 257.0 F 20.1
28:00 266.8 F 26.3
30:00 273.1 F 30.6

STERILIZING
mm:ss Degrees PSI
0:00 271.9 F 30.2
1:00 272.0 F 30.2
2:00 271.8 F 30.2
3:00 272.0 F 30.1
4:00 272.4 F 30.2

Min 271.7 F 30.1
Max 272.6 F 30.7

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 194.6 F 0.0
5:00 199.8 F 0.0
10:00 205.2 F 0.0
15:00 219.1 F 0.0
20:00 208.2 F 0.0
25:00 214.1 F 0.0
30:00 211.7 F 0.0
Drying Complete

FILLING: 0:53
HEATING: 30:20
STERILIZING: 4:00
VENTING: 3:49
DRYING: 30:00
TOTAL CYCLE: 01:09:02
CYCLE COMPLETE

03/19/2025 02:54 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____