
Midmark M11 - v1.0.5
Total Cycles: 2250

Name: RITTER2
Sterilizer ID: V2313228
04/13/2023 03:45 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 197.0 F 0.0
2:00 188.2 F 3.9
4:00 206.7 F 9.3
6:00 225.4 F 7.8
8:00 235.1 F 11.3
10:00 235.0 F 9.2
12:00 238.1 F 9.6
14:00 248.7 F 15.1
16:00 260.9 F 22.0
18:00 272.0 F 29.4

STERILIZING
mm:ss Degrees PSI
0:00 271.1 F 29.0
1:00 272.3 F 29.5
2:00 271.0 F 29.1
3:00 272.4 F 29.6
4:00 272.1 F 29.4

Min 271.0 F 28.9
Max 273.0 F 30.1

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 203.2 F 0.0
5:00 228.8 F 0.0
10:00 236.1 F 0.0
15:00 231.3 F 0.0
20:00 227.7 F 0.0
25:00 217.8 F 0.0
30:00 227.2 F 0.0
Drying Complete

FILLING: 2:30
HEATING: 18:31
STERILIZING: 4:00
VENTING: 3:16
DRYING: 30:00
TOTAL CYCLE: 00:58:16
CYCLE COMPLETE

04/13/2023 04:43 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____