
Midmark M11 - v1.0.0
Total Cycles: 680

Name: MIDMARK
Sterilizer ID: V2178198
02/19/2021 09:11 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 185.4 F 0.0
2:00 184.7 F 2.4
4:00 188.4 F 7.0
6:00 211.3 F 13.2
8:00 223.3 F 8.3
10:00 236.4 F 10.4
12:00 240.7 F 12.0
14:00 240.1 F 10.5
16:00 254.7 F 17.7
18:00 266.9 F 25.1

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 28.5
1:00 272.4 F 28.9
2:00 272.0 F 29.1
3:00 271.9 F 28.6
4:00 272.7 F 29.3

Min 271.2 F 28.3
Max 272.9 F 29.3

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 203.9 F 0.0
5:00 236.4 F 0.0
10:00 227.0 F 0.0
15:00 228.0 F 0.0
20:00 230.3 F 0.0
25:00 229.4 F 0.0
30:00 226.4 F 0.0

Drying Complete

FILLING: 1:39
HEATING: 19:27
STERILIZING: 4:00
VENTING: 2:59
DRYING: 30:00
TOTAL CYCLE: 00:58:04
CYCLE COMPLETE

02/19/2021 10:09 AM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____