
Midmark M11 - v1.0.0
Total Cycles: 1178

Name: RITTER 1
Sterilizer ID: V2178198
09/16/2021 08:22 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 169.2 F 0.0
2:00 173.3 F 2.3
4:00 186.5 F 6.1
6:00 209.1 F 11.2
8:00 219.5 F 5.5
10:00 233.4 F 10.8
12:00 231.6 F 7.5
14:00 244.7 F 12.5
16:00 244.4 F 12.3
18:00 257.0 F 19.1
20:00 268.5 F 26.1

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 28.3
1:00 272.5 F 28.9
2:00 271.5 F 28.2
3:00 272.6 F 29.0
4:00 271.6 F 28.2

Min 271.4 F 28.1
Max 272.8 F 29.1

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 200.1 F 0.0
5:00 209.7 F 0.0
10:00 222.1 F 0.0
15:00 222.1 F 0.0
20:00 219.8 F 0.0
25:00 216.2 F 0.0
30:00 181.3 F 0.0
Drying Complete

FILLING: 1:17
HEATING: 21:13
STERILIZING: 4:00
VENTING: 3:02
DRYING: 30:00
TOTAL CYCLE: 00:59:32
CYCLE COMPLETE

09/16/2021 09:21 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____