
Midmark M11 - v1.0.0
Total Cycles: 1303

Name: RITTER 1
Sterilizer ID: V2178198
10/26/2021 02:11 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 191.5 F 0.0
2:00 202.1 F 2.5
4:00 213.4 F 6.9
6:00 222.1 F 11.9
8:00 228.6 F 6.9
10:00 241.5 F 11.8
12:00 236.3 F 8.9
14:00 239.8 F 10.1
16:00 247.0 F 13.6
18:00 258.7 F 19.9
20:00 269.1 F 26.5

STERILIZING
mm:ss Degrees PSI
0:00 272.5 F 28.8
1:00 272.4 F 29.0
2:00 272.5 F 29.1
3:00 272.4 F 29.2
4:00 272.4 F 29.1

Min 272.1 F 28.7
Max 272.9 F 29.3

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 204.2 F 0.0
5:00 221.7 F 0.0
10:00 223.4 F 0.0
15:00 218.0 F 0.0
20:00 214.5 F 0.0
25:00 211.8 F 0.0
30:00 208.8 F 0.0
Drying Complete

FILLING: 1:54
HEATING: 21:46
STERILIZING: 4:00
VENTING: 3:23
DRYING: 30:00
TOTAL CYCLE: 01:01:03
CYCLE COMPLETE

10/26/2021 03:12 PM
ACCEPTED: X
REJECTED:
Operator: U01