

-----  
Midmark M11 - v1.0.5  
Total Cycles: 3610

Name: RITTER2  
Sterilizer ID: V2313228  
05/13/2024 10:36 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 162.5 F 0.0  
2:00 157.6 F 1.9  
4:00 159.0 F 4.7  
6:00 181.8 F 8.3  
8:00 197.7 F 12.6  
10:00 217.7 F 6.1  
12:00 231.4 F 10.3  
14:00 231.8 F 8.6  
16:00 239.9 F 11.9  
18:00 238.0 F 10.1  
20:00 250.3 F 16.0  
22:00 260.7 F 22.0  
24:00 270.4 F 28.3

STERILIZING  
mm:ss Degrees PSI  
0:00 271.6 F 29.1  
1:00 272.1 F 30.0  
2:00 272.1 F 29.5  
3:00 272.7 F 30.1  
4:00 271.6 F 29.7

Min 270.9 F 28.9  
Max 273.2 F 30.2

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 205.5 F 0.0  
5:00 227.2 F 0.0  
10:00 198.7 F 0.0  
15:00 196.2 F 0.0  
20:00 223.1 F 0.0  
25:00 222.7 F 0.0  
30:00 220.3 F 0.0  
Drying Complete

FILLING: 1:08  
HEATING: 24:48  
STERILIZING: 4:00  
VENTING: 3:36  
DRYING: 30:00  
TOTAL CYCLE: 01:03:33  
CYCLE COMPLETE

05/13/2024 11:39 AM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_