
Midmark M11 - v1.0.5
Total Cycles: 3608

Name: RITTER2
Sterilizer ID: V2313228
05/10/2024 03:12 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 195.7 F 0.0
2:00 194.3 F 3.9
4:00 198.2 F 9.8
6:00 218.4 F 6.1
8:00 232.9 F 11.5
10:00 234.5 F 9.5
12:00 236.6 F 9.0
14:00 250.0 F 15.8
16:00 262.8 F 23.2
18:00 271.6 F 28.8

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 28.9
1:00 272.2 F 29.2
2:00 272.2 F 29.7
3:00 271.8 F 28.9
4:00 272.7 F 29.7

Min 271.4 F 28.8
Max 272.7 F 29.7

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 208.1 F 0.0
5:00 205.7 F 0.0
10:00 204.6 F 0.0
15:00 206.5 F 0.0
20:00 202.9 F 0.0
25:00 201.6 F 0.0
30:00 213.9 F 0.0
Drying Complete

FILLING: 2:24
HEATING: 18:03
STERILIZING: 4:00
VENTING: 3:08
DRYING: 30:00
TOTAL CYCLE: 00:57:35
CYCLE COMPLETE

05/10/2024 04:09 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____