
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.5uS / 0.3ppm
CYCLE NUMBER 004361
13:56 14/09/2022
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:30
PRESSURIZING 2:32
STERILIZING 3:34
136.5 C 223kPa 3:34
136.4 C 225kPa 4:05
136.4 C 222kPa 4:35
136.2 C 221kPa 5:05
136.4 C 225kPa 5:35
136.4 C 225kPa 6:05
136.0 C 218kPa 6:35
136.4 C 223kPa 7:05
136.4 C 223kPa 7:05
VENTING 7:05
AIR DRYING 8:17
CYCLE COMPLETE 8:45

Digital Signature #
F84D9B9CD6BBF8F3
