
Midmark M11 - v1.0.0
Total Cycles: 2374

Name: RITTER 1
Sterilizer ID: V2178198
09/27/2022 07:04 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 78.8 F 0.0
2:00 103.3 F 0.0
4:00 114.4 F 0.0
6:00 140.8 F 0.2
8:00 157.7 F 1.0
10:00 181.6 F 4.3
12:00 201.8 F 7.7
14:00 215.5 F 11.2
16:00 224.7 F 5.8
18:00 232.4 F 9.6
20:00 237.2 F 10.6
22:00 240.2 F 11.3
24:00 236.8 F 9.1
26:00 250.0 F 15.4
28:00 261.1 F 21.7
30:00 271.6 F 28.4

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 28.8
1:00 272.4 F 29.0
2:00 272.2 F 29.3
3:00 271.9 F 28.7
4:00 272.5 F 29.4

Min 271.4 F 28.6
Max 272.7 F 29.4

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 183.6 F 0.0
5:00 171.0 F 0.0
10:00 160.7 F 0.0
15:00 152.1 F 0.0
20:00 149.4 F 0.0
25:00 146.3 F 0.0
30:00 143.1 F 0.0
Drying Complete

FILLING: 1:00
HEATING: 31:24
STERILIZING: 4:00
VENTING: 3:14
DRYING: 30:00
TOTAL CYCLE: 01:09:38
CYCLE COMPLETE

09/27/2022 08:13 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____